



प्रदेश सरकार
प्रदेश २
सामाजिक विकास मन्त्रालय
स्वास्थ्य निर्देशनालय

जिल्ला अस्पताल सिरहा

विषय शिलबन्दी कोटेशन सम्बन्धमा

प्रथम पटक प्रकाशित मिति २०७७/०२/२६

उपरोक्त सम्बन्धमा COVID 19 को विश्वव्यापी महामारी ले गर्दा दिन दिनै COVID 19 विरामीको संख्या बढ्दै गइरहेकोले यस अस्पताल द्वारा संचालित Isolation ward मा Hospital bed, Mattress र Pillow तत्काल खरिद नगर्दा सार्वजनिक सुरक्षा हित तथा सामुदायिक स्वास्थ्यमा पर्ने संकटलाई मध्यनजर गरि सार्वजनिक खरिद ऐन २०६३ तथा सार्वजनिक नियमावली २०६४ मा भएको विशेष परिस्थितिमा खरिद गर्ने सम्बन्धी व्यवस्था अनुसार शिलबन्दी बोलपत्रको माध्यमबाट खरिद गर्नु पर्ने भएकोले, यस सम्बन्धी विस्तृत स्पेसिफिकेशन सामाजिक विकास मन्त्रालय प्रदेश नं. २ को वेबसाईट www.mosd.p2.gov.np तथा यस अस्पतालको www.sirahahospital.gov.np र सुचनापाटीमा टाँस गरिएकोछ । दरभाउपत्र खरिद गर्दा कम्पनी/फर्म/सप्लायर्स दर्ता प्रमाण पत्र, कर चुक्ताको प्रमाणपत्र आ.व.२०७५/०७६, नविकरणको प्रतिलिपि पेश गरि नगर्दै रु १०००। तिर २०७७/०२/३२ गते भित्र यस अस्पताल बाट खरिद गर्न सकिनेछ । २०७७/०३/०१ गते दिनको १२ :०० बजे भित्र शिलबन्दी बोलपत्र दर्ता गराइ सोहि दिनको २:०० बजे प्रस्तावदाता वा निजहरुको प्रतिनिधि र यस कार्यालयका कर्मचारीहरुको रोहवरमा खोलिनेछ । जमानत वापतको धरौटी रकम रा.वा.बैंक सिरहा प्रदेश को.ले.नि.का. को नाममा रहेको खाता नं.२१३०५०२००००००० र जिल्ला अस्पतालको कार्यालय कोड ३५०००१६०१ मा तर्पिसल बमोजि रकम जम्मा गरी सककलै भौचर पेश गर्नु पर्ने छ ।

सुचना नं.	कामको विवरण	लगत अनुमान भ्याट बाहेक	धरौटी करम	कैफियत
७/०७६/०७७	Patient Bed		३६०००	
७/०७६/०७७	Bed Mattress		९०००	
७/०७६/०७७	Pillow		३०००	

डा. नागेन्द्र प्रसाद यादव
मेडिकल सुपरिटेन्डेन्ट

मेडिकल सुपरिटेन्डेन्ट



Siraha District Hospital

Siraha, Nepal

Technical specification of Plain Hospital Bed

S.N.	Purchaser's Specifications	Bidder's Remarks
	Plain Hospital Bed	
	Manufacturer	
	Brand	
	Type / Model	
	Country of Origin	
1	Technical Specifications	
1.1	Framework made of rectangle tubular framework	
1.2	Top made of perforated CRCA MS sheet	
1.3	Tubular head and foot bows of unequal height	
1.4	Size: 2100×900×550 mm	
1.5	Location for Four IV rod.	
1.6	Finish: pre - treated epoxy powdered coated	
3	Warranty	
3.1	Comprehensive warranty of anti corrosion for 1 year after acceptance.	
4.	Standards and Safety Requirements	
4.1	Must submit ISO 9001 or ISO 13485:2003/AC: 2007 or CE Certified	

Dr. Shailendra K. Paudyal
NEC no: 45 "A" Biomedical
District Hospital Siraha

[Signature]

[Signature]
मेडिकल सुपरिण्टेण्डण्ट

Siraha District Hospital

Siraha, Nepal



Technical specification of Bed Mattress

S.N.	Hospital Bed Mattress	Bidder's Remarks
	Hospital Bed Mattress	
	Manufacturer	
	Brand	
	Type/Model	
	Country of Origin	
1	Description of Function	
1.1	Mattress is to provide a comfortable platform to rest or sleep upon the bed.	
2	Operational Requirements	
2.1	Rubberized coir mattress for hospital bed.	
3	System Configuration	
3.1	Rubberized coir hospital bed mattress	
4	Technical Specifications	
4.1	The mattress negotiable with hospital beds having an overall approximate dimension of not less than 1980mm length, 910mm width.	
4.2	The mattress shall have thickness of at least 4 inches..	
4.3	The corners of the mattress shall be round shaped.	
4.4	The mattress shall be fire retardant, antibacterial treated.	
4.5	The weight bearing capacity (patient load) of the mattress shall be more than 100kg.	
4.6	The color of the mattress cover shall be finalized during the contract agreement.	
5	Standards and Safety Requirements	
5.1	Must submit ISO 9001 or ISO 13485:2003/AC: 2007 or CE Certify	
6	Warranty	
6.1	Warranty for 1 year after acceptance.	

Dr. Shailendra K. Yadav
MEC :- 45 "A" Biomedical:
District Hospital Siraha

(Signature)

88
मेडिकल सुपरिण्टेण्डण्ट



Siraha District Hospital

Siraha, Nepal

Technical specification of Hospital Bed Pillow

S.N.	Hospital Bed Pillow	Bidder's Remarks
	Hospital Bed Pillow	
	Manufacturer	
	Brand	
	Type/Model	
	Country of Origin	
1.	Technical Specifications	
4.1	Size: Approx. 17" x 26"	
4.2	Design: plain	
4.3	Material: Polyester	
4.4	Outer material: Cotton	
4.5	Shape: Rectangular	
4.6	Light weight	
5	Standards and Safety Requirements	
5.1	Must submit ISO 9001 or ISO 13485:2003/AC: 2007 or CE Certify	
6	Warranty	
6.1	Warranty for 1 year after acceptance.	

Dr. Shaktendra Pr. Yadav
MEC : 45 "A" Biomedical
BISMA Hospital Siraha

मेडिकल सुपरिण्टेण्डण्ट